

# Verification of Compliance



No. EC.1282.0H140325.PSI0999

Test Report no. XMT0201401012P/LVD, XMT0201401012P/EMC

**Certificate's Holder:** Preekem Scientific Instruments Co., Ltd.  
 No. 2 Building, NO. 4299 Jindu Road, Xinzhuang  
 Industrial Area, Shanghai, China

**Product:** Microwave Digestion System  
**Model(s):** WX-6000

**Directives:** 2006/95/EC Low Voltage  
 2004/108/EC Electromagnetic Compatibility

**Standards:** EN 61010-1:2010, EN 61000-6-1:2007,  
 EN 61000-6-3:2007, EN 61000-3-2:2006+A2:2009,  
 EN 61000-3-3:2013, BS EN 61326-1:2013,  
 BS EN 61326-2-1:2013

**Remark:** This Verification of Compliance has been issued on a voluntary basis. ECM confirms that a Technical Construction File (TCF) is existent for the above listed product(s). The TCF satisfactorily covers the essential requirements of the above listed Directive(s).

Other relevant Directives have to be observed in case they are applicable.

This Document is only valid for the equipment and configuration described and in conjunction with the TCF detailed above. Whereas the Manufacturer is responsible of the certification of the product(s) and not exempted to perform all the necessary activities before placing the product(s) on the market.

The Manufacturer is also responsible of the internal production control to ensure the product(s) are in compliance with the essential requirements of the above mentioned Directive(s).

This certificate can be checked for validity at [www.entecerma.org](http://www.entecerma.org)

Date of issue MARCH 2014

Certification Chief Manager  
 Tim Mahan



Expiry date MARCH 2019

Certification Deputy Manager  
 Viola Miller

